

This declaration of conformity is issued under the sole responsibility of

Rosemount Inc.
6021 Innovation Blvd
Shakopee, MN 55379
USA

that the following products,

Rosemount™ 3051SMV Series Pressure Transmitters
Rosemount™ 300SMV Series Pressure Transmitters

comply with the provisions of the European Union Directives, including the latest amendments, valid at the time this declaration was signed.



November 26, 2024

(signature & date of issue)

Mark Lee |
(name)

Vice President, Quality |
(function)

Boulder, CO, USA
(place of issue)

Authorized Representative in Europe:
Emerson S.R.L., company No. J12/88/2006
Emerson 4 street, Parcul Industrial
Tetarom II, Cluj-Napoca 400638, Romania

Regulatory Compliance Shared Services Department
Email: europaeproductcompliance@emerson.com Phone: +40 374 132 035

PED Notified Body:

DNV GL Business Assurance Italia S.r.l [Notified Body Number: 0496]
Via Energy Park
14, N-20871, Bimercate (MB)
Italy

ATEX Notified Bodies for EU Type Examination Certificates:

DEKRA Certification B.V. [Notified Body Number: 0344]
Utrechtseweg 310, Postbus 5185
6802 ED Arnhem
Netherlands

SGS Fimko Oy [Notified Body Number: 0598]
Takomotie 8
FI-00380 Helsinki
Finland

ATEX Notified Body for Quality Assurance:

SGS Fimko Oy [Notified Body Number: 0598]
Takomotie 8
FI-00380 Helsinki
Finland

EMC Directive (2014/30/EU)

Harmonized Standards:

EN 61326-1:2013

EN 61326-2-3: 2013

PED Directive (2014/68/EU)

Rosemount 3051SMV with Static Pressure Range 4, Differential Pressure Range 2-5; Static Pressure 5, Differential Range 2-4; and No Static Pressure, Differential Pressure Range 2-5 Pressure Transmitter (also with P0 & P9 options):

QS Certificate of Assessment –

Certificate No. 12695-2018-CE-ACCREDIA

Module H Conformity Assessment

All other Rosemount 3051SMV and 300SMV Pressure Transmitters:

Sound Engineering Practice

Transmitter Attachments: Diaphragm Seal – Process Flange – Manifold:

Sound Engineering Practice

Model 3051SFx Flowmeter Transmitters:

See DSI 1000 Declaration of Conformity for 3051SF Series Flowmeter Information

ATEX Directive (2014/34/EU)

BAS08ATEX0064X – intrinsic safety

Group II Category 1G

Ex ia IIC T4 Ga

Harmonized Standards Used:

EN IEC 60079-0:2018

EN 60079-11:2012

BAS08ATEX0065X – type “n” & increased safety

Group II Category 3G

Ex nA IIC T4 Gc

Ex ec IIC T4 Gc

Harmonized Standards Used:

EN IEC 60079-0:2018

EN 60079-15:2010

EN IEC 60079-7:2015+A1:2018

BAS01ATEX1374X – dust ignition protection by enclosure

Group II Category 1D

Ex ta IIIC T₅₀₀105°C Da

Harmonized Standards Used:

EN IEC 60079-0:2018

EN 60079-31:2014

KEMA00ATEX2143X – flameproof enclosures

Equipment Group II Category 1/2 G

Ex db IIC T6...T4 Ga/Gb

Harmonized Standards Used:

EN 60079-0: 2018

EN 60079-1:2014

EN 60079-26:2015