



EU DECLARATION OF CONFORMITY



This declaration of conformity is issued under the sole responsibility of

Rosemount Inc.
6021 Innovation Blvd
Shakopee, MN 55379
USA

that the following products,

Rosemount™ Oxygen Analyzers, Models 6888A & 6888C

comply with the provisions of the European Union Directives, including the latest amendments, valid at the time this declaration was signed.

June 13, 2024

(signature & date of issue)

Mark Lee	Vice President, Quality	Boulder, CO, USA
(name)	(function)	(place of issue)

Authorized Representative in Europe:
Emerson S.R.L., company No. J12/88/2006
Emerson 4 street, Parcul Industrial
Tetarom II, Cluj-Napoca 400638, Romania

Regulatory Compliance Shared Services Department
Email: eupeproductcompliance@emerson.com Phone: +40 374 132 035

ATEX Notified Body for EU Type Examination Certificate:
CSA Group Netherlands B.V. [Notified Body Number: 2813]
Utrechtseweg 310
6812 AR ARNHEM
Netherlands

ATEX Notified Body for Quality Assurance:
SGS Fimko Oy [Notified Body Number: 0598]
Takomotie 8
FI-00380 Helsinki
Finland

EMC Directive (2014/30/EU)

Harmonized Standards:
EN 61326-1:2013

Low Voltage Directive (2014/35/EU)

Harmonized Standards:
EN 61010-1:2010

ATEX Directive (2014/34/EU)

(Only valid for Model 6888C)

Sira14ATEX1031X – Flameproof

Equipment Group II 2 G
Ex db IIB+H2 T3 Gb; IP66
-40°C ≤ Ta ≤ +70°C Autocal Enclosure and Probe Assembly
-40°C ≤ Ta ≤ +90°C Standard Enclosure and Probe Assembly
eq. “DR Probe”

Harmonized Standards:
EN IEC 60079-0:2018
EN 60079-1:2014