


This declaration of conformity is issued under the sole responsibility of

Rosemount Inc.
6021 Innovation Blvd
Shakopee, MN 55379
USA

that the following products,

Oxymitter™ 4000, Models OXT4A, OXT4C, OXT4CNF, 6A00094G02, 6A00094G06
Oxymitter™ DR, Models OXT4ADR, OXT4CDR & OXT4CDRNF

comply with the provisions of the European Union Directives, including the latest amendments, valid at the time this declaration was signed.



JUNE 13, 2024

(signature & date of issue)

Mark Lee
(name)

Vice President, Quality
(function)

Boulder, CO, USA
(place of issue)

Authorized Representative in Europe:
Emerson S.R.L., company No. J12/88/2006
Emerson 4 street, Parcul Industrial
Tetarom II, Cluj-Napoca 400638, Romania

Regulatory Compliance Shared Services Department
Email: europaeproductcompliance@emerson.com Phone: +40 374 132 035

ATEX Notified Bodies for EU Type Examination Certificates:

CSA Group Netherlands B.V. [Notified Body Number: 2813]
Utrechtseweg 310
6812 AR ARNHEM
Netherlands

ATEX Notified Body for Quality Assurance:

SGS Fimko Oy [Notified Body Number: 0598]
Takomotie 8
FI-00380 Helsinki
Finland



EU DECLARATION OF CONFORMITY



EMC Directive (2014/30/EU)

Harmonized Standards:
EN 61326-1:2013

Low Voltage Directive (2014/35/EU)

Harmonized Standards:
EN 61010-1:2010

ATEX Directive (2014/34/EU)

(Only valid for Models OXT4C, OXT4CNF, OXT4CDR, OXT4CDRNF, 6A00094G02 & 6A00094G06)

Sira14ATEX1201X – Flameproof

Integral Assembly or Remote Probe

Equipment Group II 2 G

Ex db IIB+H2 T4 Gb (-40°C ≤ Ta ≤ +70°C)

Integral Assembly or Remote Probe without flame arrestor

Equipment Group II 2/- G

Ex db IIB+H2 T4 Gb/- (-40°C ≤ Ta ≤ +70°C)

Split Architecture Remote Electronics

Equipment Group II 2 G

Ex db eb IIB+H2 T5 Gb (-40°C ≤ Ta ≤ +70°C)

Harmonized Standards:

EN IEC 60079-0:2018

EN 60079-1:2014

EN 60079-7:2015+A1:2018