

EC DECLARATION OF CONFORMITY

Document number: RAE/X-STREAM XE DC-E1

Date: August 2009

We,

Emerson Process Management GmbH & Co. OHG

located at

Industriestrasse 1, D-63594 Hasselroth, Germany

declare under our sole responsibility that our gas analyzer, type

X-STREAM XE

to which this declaration relates is in conformity with the provisions of:

2004/108/EC EMC Directive

with the application of the harmonized standard:

EN 61326-1:2006 Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements

97/23/EC Pressure Equipment Directive

This analyzer has been designed and manufactured considering article 3, paragraph 3 of the above mentioned directive and therefore CE marking does not refer to this directive.

This document covers all ½ 19" X-STREAM XE gas analyzer variations with DC power supply.

Hasselroth, August 2009


.....
(Signature)

Andy Kemish

(Name)

VP Rosemount Analytical Europe

(Function name)

ROSEMOUNT®
Analytical

This declaration confirms the compliance with announced directives but does not include the assurance of properties.
The safety and installation instructions of the documentation have to be followed.


EMERSON
Process Management