Product Certifications 00880-0200-4480, Rev AE December 2024

Rosemount[™] 1408A Level and Flow Transmitter

Non-Contacting Radar





ROSEMOUNT

1 Product certifications

Rev 2.23

Note

The most recent revision can be found at Emerson.com/Rosemount1408A.

1.1 European directive information

A copy of the EU Declaration of Conformity can be found at the end of the document. The most recent revision of the EU Declaration of Conformity can be found at <u>Emerson.com/Rosemount</u>.

1.2 Ordinary location certification

As standard, the transmitter has been examined and tested to determine that the design meets the basic electrical, mechanical, and fire protection requirements by a nationally recognized test laboratory (NRTL) as accredited by the Federal Occupational Safety and Health Administration (OSHA).

Certificate	80031621			
Standards	CAN/CSA-C22.2 No. 61010-1-12,			
	UL Std. No. 61010-1			

The device may only be powered by a power supply unit with a limited energy electric circuit max. 30 Vdc (IO-Link) / 35 Vdc (HART[®]) output in accordance with CAN/CSA-C22.2 No. 61010-1-12 / UL Std. No. 61010-1 (3rd Edition) chapter 6.3.1/6.3.2 and 9.4 or class 2 according to CSA 223/UL 1310.

1.3 Environmental conditions

Table 1-1: Environmental Conditions (Ordinary Location and Low Voltage Directive (LVD))

Туре	Description		
Location	Indoor or outdoor use, wet ⁽¹⁾		
Maximum altitude	6562 ft. (2000 m)		
Ambient temperature	-40 to 176 °F (-40 to 80 °C)		
Installation category	DC supplied		
Electrical supply for IO-Link device	18-30 Vdc, 3.6 W		
Electrical supply for HART [®] device	14-35 Vdc, 0.8 W (35Vdc @ 22.5mA ⁽²⁾)		
Mains supply voltage fluctuations for IO-Link device	Safe at 18-30 Vdc ±10%		
Mains supply voltage fluctuations for HART device	Safe at 14-35 Vdc ±10%		
Pollution degree	2		

- (1) Outdoor use and wet location not part of Ordinary Location certification.
- (2) Nominal power consumption is 0.54 W (24Vdc @ 22.5 mA).

1.4 Telecommunication compliance

Rosemount 1408A, with the antenna extension fitted, is a device for measurement of level in the open air or in an enclosure. When used for measurement in an enclosure (i.e metallic or reinforced concrete or fiberglass tanks, or similar enclosure structures made of comparable attenuating material), the antenna extension may be omitted.

Hardware Version Identification Number (HVIN) is 1408L for the IOlink device and 1408LH for the HART device.

Measurement principle

Frequency Modulated Continuous Wave (FMCW), 80 GHz

Maximum output power

3 dBm (2 mW)

Frequency range

77 to 81 GHz

76-77 GHz in applicable countries, contact Emerson for details.

1.5 FCC

Note: This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

FCC ID K8C1408L

1.6 IC

This device complies with Industry Canada's licence-exempt RSS standard. Operation is subject to the following conditions:

- 1. This device may not cause harmful interference.
- 2. This device must accept any interference received, including interference that may cause undesired operation.
- 3. The installation of the LPR device shall be done by trained installers in strict compliance with the manufacturer's instructions.
- 4. The use of this device is on a "no-interference, no-protection" basis. That is, the user shall accept operations of high-powered radar in the same frequency band which may interfere with or damage this device. However, devices found to interfere with primary licensing operations will be required to be removed at the user's expense.
- 5. The installer/user of this device shall ensure that it is at least 10 km from the Dominion Astrophysical Radio Observatory (DRAO) near Penticton, British Columbia. The coordinates of the DRAO are latitude 49°19′15″N and longitude 119°37′12″ W. For devices not meeting this 10 km separation (e.g., those in

the Okanagan Valley, British Columbia,) the installer/user must coordinate with, and obtain the written concurrence of, the Director of the DRAO before the equipment can be installed or operated. The Director of the DRAO may be contacted at 250-497-2300 (tel.) or 250-497-2355 (fax). (Alternatively, the Manager, Regulatory Standards, Industry Canada, may be contacted.)

Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux conditions suivantes:

- 1. l'appareil ne doit pas produire de brouillage.
- l'appareil doit accepter tout brouillage radioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.
- L'installation d'un dispositif LPR doit être effectuée par des installateurs qualifiés, en pleine conformité avec les instructions du fabricant.
- 4. Ce dispositif ne peut être exploité qu'en régime de nonbrouillage et de non-protection, c'est-à-dire que l'utilisateur doit accepter que des radars de haute puissance de la même bande de fréquences puissent brouiller ce dispositif ou même l'endommager. D'autre part, les capteurs de niveau qui perturbent une exploitation autorisée par licence de fonctionnement principal doivent être enlevés aux frais de leur utilisateur.
- 5. La personne qui installe/utilise ce capteur de niveau doit s'assurer qu'il se trouve à au moins 10 km de l'Observatoire fédéral de radioastrophysique (OFR) de Penticton en Colombie-Britannique. Les coordonnées de l'OFR sont: latitude N 49° 19′ 15″, longitude O 119° 37′ 12″. La personne qui installe/utilise un dispositif ne pouvant respecter cette distance de 10 km (p. ex. dans la vallée de l'Okanagan [Colombie-Britannique]) doit se concerter avec le directeur de l'OFR afin d'obtenir de sa part une autorisation écrite avant que l'équipement ne puisse être installé ou mis en marche. Le directeur de l'OFR peut être contacté au 250-497-2300 (tél.) ou au 250-497-2355 (fax). (Le Directeur des Normes réglementaires d'Industrie Canada peut également être contacté).

Certificate 2827A-1408L

1.7 Radio Equipment Directive (RED) 2014/53/EU

D Open air installations

Rosemount 1408A, when fitted with the antenna extension, complies with ETSI EN 302 729 and EN 62479.

For open air installations, restrictions exist in several EU and EFTA member states in the proximity of Radio Astronomy sites. Locations of Radio Astronomy sites completely or partially using the same frequency band as Rosemount 1408A (77 to 81 GHz) vary over time. Therefore, the installer and the user of Rosemount 1408A is obligated to check how their installation location is affected.

The current list of Radio Astronomy sites including their corresponding operating frequencies may be found at <u>www.craf.eu</u>.

The following restrictions apply:

- Install at a separation distance of >4 km from Radio Astronomy sites measuring at 77-81 GHz, unless a special authorization has been provided by the responsible National regulatory authority.
- Between 4 km to 40 km around any Radio Astronomy site, the LPR antenna height shall not exceed 15 m above ground.

Closed tanks

Rosemount 1408A without the antenna extension complies with ETSI EN 302 372 and EN 62479.

The device must be installed in closed tanks (metal, reinforced concrete tanks or similar enclosure structures made of comparable attenuating material). Install according to requirements in ETSI EN 302 372 (Annex E).

Performance under the influence of an interferer signal

For the receiver test that covers the influence of an interferer signal to the device, the performance criterion has at least the following level of performance according to ETSI TS 103 361 [6].

- Performance criterion: measurement value variation Δd over time during a distance measurement
- Level of performance: $\Delta d \le \pm 2 \text{ mm}$

1.8 Radio/EMC Australia and New Zealand

Rosemount 1408A complies with the requirements of the relevant ACMA Standards made under the Radiocommunications Act 1992 and the Telecommunications Act 1997 and the relevant Standards made under The New Zealand Radio Communication Act 1989.

In New Zealand, Rosemount 1408A must be installed in closed tanks (metal, reinforced concrete tanks or similar enclosure structures made of comparable attenuating material).

- 1.9 Other radio approvals
- 1.9.1 Argentina



H-30153

1.9.2 Brazil

This equipment is not entitled to protection against harmful interference and may not cause interference with properly authorized systems.

This product is not suitable for use in residential environments, as it may cause electromagnetic interference that requires the user to take necessary actions to minimize such interference.

The product installation must be located at least 4 km away from the radio astronomy stations.

The installation height of the product must be less than 15 meters, when installed at a distance of up to 40 km from the radio astronomy stations. This model can be used in outdoor and indoor environments.

1.9.3 Malaysia



1.9.4 Republic of Korea

Registration number R-R-

- R-R-Rtr-1408
- 1.9.5 Singapore

Complies with IMDA standards DA100927

1.9.6 South Africa



1.9.7 Taiwan



Without permission granted by the NCC, any company, enterprise, or user is not allowed to change frequency, enhance transmitting power or alter original characteristic as well as performance to an approved low power radio-frequency device. The low power radio-frequency devices shall not influence aircraft security and interfere legal communications; If found, the user shall cease operating immediately until no interference is achieved. The said legal communications means radio communications is operated in compliance with the Telecommunications Management Act. The low power radio-frequency devices must accept any interference from legal communications or ISM radiowave radiated devices.

1.9.8 Thailand

This telecommunication equipment conforms to NTC technical requirements.

เครื่องโทรคมนาคมและอุปกรณ์นี้มีความสอดคล้องตามมาตรฐานหรือข้อกำหนด ทางเทคนิค ของ กสทช

1.9.9 Other national spectrum approvals

Radio devices usually require certification to ensure they adhere to regulations regarding the use of radio frequency (RF) spectrum. Many countries require this type of product certification.

Emerson is working with governmental agencies around the world to supply fully compliant products and remove the risk of violating country directives or laws governing radio device usage.

1.10 Canadian Registration Number (CRN)

Transmitter without adapter

Registrations Alberta (ABSA): 0F21418.2 British Columbia (TSBC): 0F7358.1 Manitoba (ITS): 0F21418.24 New Brunswick: 0F21418.27 Newfoundland and Labrador: 0F21418.20 Northwest Territories: 0F21418.2T Nova Scotia: 0F21418.28 Nunavut: 0F21418.2N Ontario (TSSA): 0F23714.5 Prince Edward Island: 0F21418.29 Québec (RBQ): 0F05457.6 Saskatchewan (TSASK): 0F2113.3 Yukon: 0F21418.2Y

1.11 EU Declaration of Conformity

Figure 1-1: EU Declaration of Conformity





Page 2 of 2

1.12 China RoHS

召有Unina RonS官控物质超过最大浓度限值的部件型号列表									
Part Name 部件名称	Hazardous Substances / 有害物质								
	Lead 铅 (Pb)	Mercury 汞 (Hg)	Cadmium 镉 (Cd)	Hexavalent Chromium 六价铬 (Cr +6)	Polybrominated biphenyls 多溴联苯 (PBB)	Polybrominated diphenyl ethers 多溴联苯醚 (PBDE)			
Electronics Assembly 电子组件	х	ο	ο	0	0	0			
Housing Assembly 壳体组件	0	ο	ο	0	0	0			

List of Model Parts with China RoHS Concentration above MCVs 含有China RoHS 管控物质超过最大浓度限值的部件型号列表

This table is proposed in accordance with the provision of SJ/T11364

本表格系依据SJ/T11364的规定而制作.

O: Indicate that said hazardous substance in all of the homogeneous materials for this part is below the limit requirement of GB/T 26572.

O: 意为该部件的所有均质材料中该有害物质的含量均低于GB/T 26572所规定的限量要求.

X: Indicate that said hazardous substance contained in at least one of the homogeneous materials used for this part is above the limit requirement of GB/T 26572.

X: 意为在该部件所使用的所有均质材料里,至少有一类均质材料中该有害物质的含量高于GB/T 26572所规定的限量要求.

Product Certifications 00880-0200-4480, Rev. AE December 2024

For more information: Emerson.com/global

©2024 Emerson. All rights reserved.

Emerson Terms and Conditions of Sale are available upon request. The Emerson logo is a trademark and service mark of Emerson Electric Co. Rosemount is a mark of one of the Emerson family of companies. All other marks are the property of their respective owners.



ROSEMOUNT