

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Chemiluminescence Detection Gas Analyzer

Name and address of the applicant

Emerson Process Management GmbH & Co. OHG
Industriestrasse 1,
DE-63594 Hasselroth, Germany

Name and address of the manufacturer

Emerson Process Management GmbH & Co. OHG
Industriestrasse 1,
DE-63594 Hasselroth, Germany

Name and address of the factory

Emerson Process Management GmbH & Co. OHG
Industriestrasse 1,
DE-63594 Hasselroth, Germany

Note: When more than one factory, please report on page 2

 Additional Information on page 2

Ratings and principal characteristics

100-240 V~, 50/60 Hz, 2-1 A

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

X-STREAM Enhanced Series, model XECLD

Additional information (if necessary may also be reported on page 2)

Modification to cover additional PFC-Inductor to CB Certificate CA/23271/CSA, issued 2019-08-09.

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010+AMD1:2016,
including ND for CA and US

As shown in the Test Report Ref. No. which forms part of this Certificate

185562-80060741

This CB Test Certificate is issued by the National Certification Body

CSA Group
178 Rexdale Boulevard
Toronto, ON M9W 1R3 Canada

Date: 2020-12-16

Signature: Vegard Andersen